



Interaural bimodal pitch matching with two-formant vowels

Guerit, Francois Marie Louis Paul; Chalupper, Josef; Santurette, Sébastien; Arweiler, Iris; Dau, Torsten

Publication date:
2013

Document Version
Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

Citation (APA):
Guerit, F. M. L. P., Chalupper, J., Santurette, S., Arweiler, I., & Dau, T. (2013). *Interaural bimodal pitch matching with two-formant vowels*. Poster session presented at International Symposium on Auditory and Audiological Research, Nyborg, Denmark.

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Interaural bimodal pitch matching with two-formant vowels

François Guérit¹, Josef Chalupper², Sébastien Santurette¹,
Iris Arweiler², Torsten Dau¹

¹Centre for Applied Hearing Research, Technical University of Denmark

²Advanced Bionics GmbH

Introduction

For bimodal patients, with a hearing aid (HA) in one ear and a cochlear implant (CI) in the opposite ear, usually a default frequency-to-electrode map is used in the CI. This assumes that the human brain can adapt to interaural place-pitch mismatches. This "one-size-fits-all" method might be partly responsible for the large variability of individual bimodal benefit. Therefore, knowledge about the location of the electrode array is an important prerequisite for optimum fitting. Theoretically, the electrode location can be determined from CT-scans. However, these are often not available in audiological practice. Behavioral pitch matching between the two ears has also been suggested, but has been shown to be tedious and unreliable (Carlyon et al., 2010). Here, an alternative method using two-formant vowels was developed and tested.

Hypothesis




Fig.1 If the implant is perfectly fitted, the three vowels of this example should be perceived identically when presenting the second formant (F2, yellow) in either the NH or CI side. However, if there is a shift towards the base, the perceived vowel map obtained by varying F2 (dashed white rectangle) should also show a shift.


Methods

Stimuli Two-formants vowels were produced using Matlab-based Klatt synthesizer (Klatt, 1980) and mixed with consonants to form a /t/-vowel/-k/ stimulus. Noise-vocoder (Litvak et al., 2007) was simulating in the right ear a perfect fit and two different mismatches.

Procedure Subjects had to categorize (forced choice) the perceived stimuli into different vowel propositions using a Matlab GUI.

Subjects All subjects were native German speakers. Tests with normal hearing (NH) subjects were done in Denmark, tests with bimodal subjects were achieved in the ENT department of the UnfallKrankenHaus, Berlin.

NH without vocoder



Settings:

- F1 presented on the left ear
- F2 presented either on the left or right ear
- 10 repetitions per stimulus
- 4 NH subjects
- Possible choices: TUK, TÜK, TIK, TOK, TÖK, TEK, TAK, TÄK

Results:

- similar results when having F2 in the left or right ear
- if F1 is fixed at 250 Hz, TUK, TÜK and TIK are successively perceived by increasing F2
- if F1 is fixed at 400 Hz, TOK, TÖK and TEK are successively perceived by increasing F2

Fig.2 Distribution of the "TÜK" category for all subjects, scaled from 0% (white) to 100% (bright red)

NH with vocoder




Fig.3 Distribution of vowel categories for subject "NH4" when changing F2, with F1 fixed at 250 Hz in the left ear. (A): F2 presented in the left ear. (B), (C) and (D): vocoded F2 presented in the right ear. Areas show a fitted gaussian curve of vowel distribution (markers). To help the eye, center frequency of the fitted "TÜK" (solid black line) and its expected frequency from the vocoder settings (dashed line) are shown.

Settings:

- F1=250 and 400 Hz, F2 presented either on the left or vocoded on the right ear
- 5 repetitions (F2 on the left ear)/ 15 repetitions (F2 on the right ear) per stimulus
- 8 NH subjects
- Possible choices: TUK, TÜK, TIK, TOK, TÖK, TEK
- vocoder training with audiobook

Results:

- Fusion of percepts across ears was difficult but possible after audiobook training
- individual vowel maps are affected by simulated mismatches
- variability is high between subjects, especially when simulating the big mismatch




Fig. 4 Eight subjects fitted center frequencies (dark) for the vowel TUK (left) and TÖK (right). Expected center frequencies are shown in dashed gray lines.




Fig. 5 Eight subjects fitted vowel distributions, with vocoder settings 2 (top) and 3 (bottom), simulating respectively a small and big mismatch. F1=250 Hz.

Bimodal patients




Fig. 6 Six bimodal patients vowel distributions, when F2 was presented acoustically (left of each subject panel) and electrically (right).

F1=250 Hz only was presented (choices reduced to TUK, TÜK, TEK), with 10 repetitions. There is an important variability, future CT-scans comparison might help interpreting the results.

Discussion

- small shift simulation with a vocoder(≈ 500 Hz) can be estimated by looking at the distribution of three-vowels groups

- bigger shift (≈ 800 Hz) creates an important variability, probably due to range limitations and difficulty to fuse percepts

- variability in bimodal patients is likely to come from nonsensory biases arising from the very different percepts in the two ears, as shown in (Carlyon et al., 2010)

References

- Carlyon et al. (2010). Pitch comparisons between electrical stimulation of a cochlear implant and acoustic stimuli presented to a normal-hearing contralateral ear. *JARO*
- Litvak et al. (2007). Relationship between perception of spectral ripple and speech recognition in cochlear implant and vocoder listeners. *JASA*
- Klatt, D. H. (1980). Software for a cascade/parallel formant synthesizer. *JASA*